

9-1-2015

Aneurysm permeability following coil embolization: packing density and coil distribution

Juyu Chueh

University of Massachusetts Medical School, Juyu.Chueh@umassmed.edu

Srinivasan Vedantham

University of Massachusetts Medical School, Srinivasan.Vedantham@umassmed.edu

Ajay K. Wakhloo

University of Massachusetts Medical School, Ajay.Wakhloo@umassmemorial.org

See next page for additional authors

Follow this and additional works at: http://escholarship.umassmed.edu/radiology_pubs

 Part of the [Cardiovascular Diseases Commons](#), [Nervous System Diseases Commons](#), [Radiology Commons](#), and the [Surgery Commons](#)

Repository Citation

Chueh, Juyu; Vedantham, Srinivasan; Wakhloo, Ajay K.; Carniato, Sarena L.; Puri, Ajit S.; Bzura, Conrad; Coffin, Spencer; Bogdanov, Alexei A.; and Gounis, Matthew J., "Aneurysm permeability following coil embolization: packing density and coil distribution" (2015). *Radiology Publications and Presentations*. Paper 157.

http://escholarship.umassmed.edu/radiology_pubs/157

This material is brought to you by eScholarship@UMMS. It has been accepted for inclusion in Radiology Publications and Presentations by an authorized administrator of eScholarship@UMMS. For more information, please contact Lisa.Palmer@umassmed.edu.

Aneurysm permeability following coil embolization: packing density and coil distribution

Authors

Juyu Chueh, Srinivasan Vedantham, Ajay K. Wakhloo, Sarena L. Carniato, Ajit S. Puri, Conrad Bzura, Spencer Coffin, Alexei A. Bogdanov, and Matthew J. Gounis

Keywords

Aneurysm, Coil, Intervention

Comments

This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>

Creative Commons License



This work is licensed under a [Creative Commons Attribution-Noncommercial 4.0 License](http://creativecommons.org/licenses/by-nc/4.0/)

Rights and Permissions

Citation: J Neurointerv Surg. 2015 Sep;7(9):676-81. doi: 10.1136/neurintsurg-2014-011289. Epub 2014 Jul 16. [Link to article on publisher's site](#)



OPEN ACCESS

ORIGINAL RESEARCH

Aneurysm permeability following coil embolization: packing density and coil distribution

Ju-Yu Chueh,¹ Srinivasan Vedantham,¹ Ajay K Wakhloo,¹ Sarena L Carniato,² Ajit S Puri,¹ Conrad Bzura,¹ Spencer Coffin,¹ Alexei A Bogdanov Jr,¹ Matthew J Gounis¹

¹Department of Radiology, New England Center for Stroke Research, University of Massachusetts Medical School, Worcester, Massachusetts, USA
²PreClinical Science, Stryker Neurovascular, Fremont, California, USA

Correspondence to

Dr Matthew J Gounis, 55 Lake Avenue North, Worcester, MA 01655, USA; Matthew.Gounis@umassmed.edu

J-YC and SV contributed equally.

Received 22 May 2014
Revised 26 June 2014
Accepted 29 June 2014
Published Online First 16 July 2014

ABSTRACT

Background Rates of durable aneurysm occlusion following coil embolization vary widely, and a better understanding of coil mass mechanics is desired. The goal of this study is to evaluate the impact of packing density and coil uniformity on aneurysm permeability. **Methods** Aneurysm models were coiled using either Guglielmi detachable coils or Target coils. The permeability was assessed by taking the ratio of microspheres passing through the coil mass to those in the working fluid. Aneurysms containing coil masses were sectioned for image analysis to determine surface area fraction and coil uniformity.

Results All aneurysms were coiled to a packing density of at least 27%. Packing density, surface area fraction of the dome and neck, and uniformity of the dome were significantly correlated ($p < 0.05$). Hence, multivariate principal components-based partial least squares regression models were used to predict permeability. Similar loading vectors were obtained for packing and uniformity measures. Coil mass permeability was modeled better with the inclusion of packing and uniformity measures of the dome ($r^2 = 0.73$) than with packing density alone ($r^2 = 0.45$). The analysis indicates the importance of including a uniformity measure for coil distribution in the dome along with packing measures.

Conclusions A densely packed aneurysm with a high degree of coil mass uniformity will reduce permeability.

hemodynamic stress suggested that a coil mass at a given PD with low permeability may accelerate intra-aneurysmal thrombosis due to the drastic reduction in intra-aneurysmal velocities and increased residence times.⁵ A possible explanation for the variation in permeability at a given PD may be the differences in spatial distribution of the coil mass within the aneurysm.⁹

The goal of this study is to investigate the effects of both coil PD and spatial uniformity on aneurysm permeability. We have devised a repeatable in vitro assessment tool to objectively measure the coil uniformity in addition to the standard benchmark of PD. We hypothesized that coil permeability is influenced by a combination of PD and uniformity.

METHODS**Aneurysm model and coiling procedure**

Patient-specific A1 silicone aneurysms (5 mm dome, 3 mm neck, 2 mm parent vessel diameter), shown in [figure 1](#), were prepared using a previously described technique.¹⁰ We incorporated a 1.5 mm channel emanating from the aneurysm dome for permeability assessment (blue arrow in [figure 1](#)). Prior to aneurysm coiling, the aneurysm volume was calculated by measuring the volume of water needed to fill the aneurysm sac.

Silicone aneurysm models were packed by using Guglielmi detachable coils (GDC, $n = 8$) or Target coils ($n = 8$) (Stryker Neurovascular, Fremont, California, USA). Using an Excelsior SL10 (Stryker Neurovascular), the GDC group was framed with a 5×9 mm GDC 360° detachable coil and finished using only GDC detachable coils. In the Target group, aneurysms were coiled using the standard Target 5×15 mm framing coils and were filled and finished with Target detachable and Target nano coils, respectively. The actual type, size, and number of coils used for both groups were determined by the operator's preference. The coiling procedure was terminated after achieving the procedural end points—namely, complete aneurysm obliteration, catheter dislodgement from the aneurysm, or herniation of coil loops into the parent artery.¹¹

The number, size, and type of coils were recorded for each case and this information was used to calculate the PD of the aneurysm using *AngioCalc* (<http://www.angiocalc.com>). The PD was determined by dividing the coil volume by the total volume of the aneurysm.

INTRODUCTION

Endovascular coil embolization has emerged as the gold standard for treatment of intracranial aneurysms.^{1,2} The current standard of optimal endovascular treatment of intracranial aneurysms is based on achieving angiographic occlusion and high packing density (PD) in order to achieve a durable exclusion of the aneurysmal sac from the cerebral circulation.^{3,4} The goal of endovascular coil embolization is to reduce the flow of the aneurysm, which varies with coil properties such as permeability and therefore results in thrombus formation, exclusion from circulation, and a decreased risk of rupture.⁵

Studies have shown that increasing PD resulted in lower intra-aneurysmal velocities and less fluid flow across the aneurysm.^{6–8} Although an inverse relationship between PD and coil permeability has been observed, the PD of the coil mass may not be the only factor influencing its permeability.

Previous work on modeling the interactions between the coils within an aneurysm with the



Open Access
Scan to access more
free content



CrossMark

To cite: Chueh J-Y, Vedantham S, Wakhloo AK, et al. *J NeuroIntervent Surg* 2015;**7**:676–681.




Figure 1 Computer model of the A1 aneurysm. The blue arrow indicates the channel which emanates from the aneurysm dome.

Permeability assessment

Coil permeability was obtained by evaluating the ability of fluorescent microspheres to pass through the coil mass. Neutrally buoyant fluorescent microspheres of 1 μm that approximate the size of platelets were used in this study. Spheres of this size and specific gravity are often used for hemodynamic measurements *in vitro*.^{12–13} Calibration curves were established to convert the recorded fluorescent intensities into microsphere concentration.¹⁴

The fluorescent microsphere solution was prepared by mixing 1 μm (5×10^6 microspheres/mL) fluorescent microspheres (Invitrogen FluroSpheres, Eugene, Oregon, USA) in a mixture of 0.15 M NaCl, 0.05% Tween 20, and 0.02% thimerosal. The coiled aneurysm was connected to the flow loop filled with the fluorescent microsphere solution. The channel emanating from the aneurysm dome was open to allow the microspheres to pass through for collection. The fluorescent intensity/concentration of the collected sample was recorded and normalized against the intensity/concentration of the stock solution.

Aneurysm embedding and sectioning

Aneurysms were embedded in a low viscosity epoxy embedding medium, Spurr's Resin (SPI Supplies/Structure Probe, West Chester, Pennsylvania, USA). Coiled aneurysms were infiltrated with the resin and placed in low pressure to remove air trapped in the model. The model was cured, allowing the resin to polymerize. Once the model had cured, the silicone was removed from the aneurysm model. One neck and two dome sections were obtained from each aneurysm model using a low speed saw (Buehler, Lake Bluff, Illinois, USA) for a total of three sections. The neck section was cut between the aneurysm dome and parent vessel. The two dome sections were obtained from two vertical cuts perpendicular to the neck of the aneurysm on either side of the center line. Sections were mounted, polished, and rinsed with 2-propanol. Mounted aneurysm sections were imaged using a flat-bed scanner (CanoScan 9950F; Lake Success, New York, USA) and saved for image analysis.

Image analysis

Scanned images of aneurysm sections (figure 2A shows a dome section) were imported into MatLab (MathWorks, Natick, Massachusetts, USA), converted to a binary image (figure 2B), and custom fit with an elliptical mask that was truncated at the neck section (figure 2C) to provide a binary coil segmented image, $C(i, j)$. The mask, $M(i, j)$, is shown in figure 2D. The surface area fraction (SAF) was calculated as shown in Equation 1:

$$\text{SAF} = \frac{\sum_{i,j} C(i, j)}{\sum_{i,j} M(i, j)} \quad (1)$$

Spatial uniformity of binary coil mass segmented images (eg, figure 2C) was determined using the fractal-based heterogeneity measure, lacunarity, represented as Λ . Values of Λ approaching 0 imply spatially uniform distribution of coil mass with minimal gaps. Lacunarity was determined using the FracLac plugin (V2013April.b6, <http://rsbweb.nih.gov/ij/plugins/fracLac/>) for ImageJ (V1.44p, National Institutes of Health). A sliding-box technique with a minimum region of interest of 5×5 pixels was translated by one pixel along i and j directions to determine lacunarity. While various metrics could be used for assessing spatial uniformity, lacunarity was chosen as it is a well-established image analysis metric^{15–19} that has been applied in numerous studies including application in neuroimaging.²⁰ Further, the ready availability of this analysis tool would permit researchers to conduct similar investigations.

Statistical analysis

The collected data were statistically analyzed using SAS V9.3 (SAS Institute, Cary, North Carolina, USA). Effects associated with $p < 0.05$ were considered statistically significant. Paired




Figure 2 Image processing of aneurysm 6 (Target) including (A) the scanned image, (B) the binary image, (C) the binary image with elliptical mask adjusted to account for the neck section, and (D) the mask.

Table 1 Summary statistics for measured parameters

Parameters	Symbol	Mean±SD	Median (1st, 3rd quartiles)	Minimum, maximum
Coil mass permeability	P	0.924±0.071	0.953 (0.869, 0.975)	0.785, 1.0
Packing density	PD	0.343±0.044	0.347 (0.313, 0.358)	0.271, 0.455
Surface area fraction of dome section	SAF _D	0.488±0.106	0.460 (0.403, 0.584)	0.331, 0.669
Surface area fraction of neck section	SAF _N	0.345±0.111	0.338 (0.245, 0.423)	0.152, 0.539
Lacunarity of dome section	Λ _D	0.065±0.020	0.057 (0.051, 0.085)	0.037, 0.098
Lacunarity of neck section	Λ _N	0.112±0.047	0.117 (0.077, 0.126)	0.047, 0.196

t tests were used to determine if the two dome sections differed in SAF and lacunarity. Statistical tests were conducted to assess whether coil mass permeability, packing, and uniformity measures differed with coil type (GDC vs Target). It should be noted that the term ‘coil type’ used in this paper refers to all the variations between the GDC and the Target coiling systems, including but not limited to length of the framing coil, coil stiffness and configuration, and design of the coil/delivery wire junction. Correlations between PD, SAF, and lacunarity were determined. Multivariate linear regression models were used to predict coil mass permeability as the outcome, with packing and uniformity measures as predictor variables. If the predictor variables exhibited statistically significant correlation, then partial least squares regression using principal components (PC-PLSR) was used for statistical modeling with leave-one-out cross validation for extraction of factors. A linear regression model with coil mass permeability as the dependent variable and PD as the independent variable was used for comparative assessment.

RESULTS

All aneurysms were coiled to a PD of at least 27%. During aneurysm sectioning, five of the 48 sections (3 neck and 2 dome) were damaged and hence the remaining 43 (89.5%) sections (13 neck and 30 dome) were included in the analysis. For the two dome sections from the same aneurysm, the SAF and lacunarity measures were not statistically different ($p>0.19$, paired t test) and hence were averaged. Where applicable, subscripts N and D are used to denote the neck and dome sections, respectively. Summary statistics for all measured parameters are provided in [table 1](#). Despite relatively high PD, the coil masses exhibited substantial permeability for the microspheres. The paired t test indicated that SAF_N and SAF_D were statistically different ($p<0.001$), with the mean being higher for SAF_D than SAF_N ([table 1](#)). Similarly, Λ_N and Λ_D were statistically different ($p<0.001$), with the mean being higher for Λ_N than Λ_D, implying that the coil distribution at the cross-section of the neck was less uniform than in the dome. Independent t tests with the Satterthwaite method for unequal variances indicated that, compared with GDC coils, Target coils had higher PD ($p=0.026$), higher SAF_D ($p=0.002$), and reduced coil mass permeability ($p=0.006$). The SAF_N, Λ_N, and Λ_D were not statistically different between GDC and Target coils ($p>0.18$).

Statistical modeling

[Table 2](#) summarizes the Pearson correlation coefficients between the packing and uniformity measures.

All variables were significantly correlated with the exception of PD and Λ_N. Hence, statistical models using PC-PLSR were used when two or more predictors were included in the model and linear regression was used when only one predictor was included in the model.

Five statistical regression models were considered with coil mass permeability (P) as the outcome variable (Equation 2). Model A assumes that the coil mass permeability is influenced by PD alone and was obtained by linear regression. Since the permeability was statistically different between GDC and Target coils, Model B improves on Model A by accounting for the differing coil types (C_T). Model C considers all variables studied as predictors in the regression model, Model D considered the predictors from Model B and the packing and uniformity measures from the dome section, and Model E considered the predictors from Model B and the packing and uniformity measures from the neck section. Thus, Models D and E address the question whether the permeability is better predicted when packing and uniformity measures from either the dome or the neck section are included with the C_T and the PD. For all five statistical models the residuals were analyzed and found to satisfy the normality requirement ($p>0.07$, Shapiro-Wilks test). Wold’s variable importance for projection criterion, in conjunction with the scaled and centered parameter estimates, was used to assess whether any of the predictor variables in each PC-PLSR model needed to be excluded.

$$\left. \begin{aligned} \text{Model A : } P &= f(\text{PD}) \\ \text{Model B : } P &= f(\text{C}_T, \text{PD}) \\ \text{Model C : } P &= f(\text{C}_T, \text{PD}, \text{SAF}_D, \text{SAF}_N, \Lambda_D, \Lambda_N) \\ \text{Model D : } P &= f(\text{C}_T, \text{PD}, \text{SAF}_D, \Lambda_D) \\ \text{Model E : } P &= f(\text{C}_T, \text{PD}, \text{SAF}_N, \Lambda_N) \end{aligned} \right\} \quad (2)$$

Linear regression (Model A) indicated that the PD was a statistically significant predictor ($p=0.005$), confirming that the PD plays an important role in coil mass permeability. For each PLS-PCR regression (Models B–E), the magnitudes of the model effects loadings or weights are summarized in [table 3](#). For each model, the similar magnitudes of the predictors within that model indicate that all predictors within each model are equally important.

Table 2 Pearson correlation coefficient

	PD	SAF _N	SAF _D	Λ _N	Λ _D
PD	1	0.558*	0.740*	−0.375	−0.598*
SAF _N	0.558*	1	0.554*	−0.753*	−0.594*
SAF _D	0.740*	0.554*	1	−0.643*	−0.831*
Λ _N	−0.375	−0.753*	−0.643*	1	0.637*
Λ _D	−0.598*	−0.594*	−0.831*	0.637*	1

*Correlation is significant (two-tailed test) at the 0.05 level.
PD, packing density; SAF, surface area fraction.

Table 3 Magnitudes of the model effects loadings or weights

Statistical models	Coil type C_T	Packing density PD	Dome section		Neck section	
			SAF _D	Λ_D	SAF _N	Λ_N
Model B	0.620	0.481				
Model C	0.373	0.380	0.443	0.387	0.341	0.340
Model D	0.460	0.432	0.494	0.382		
Model E	0.487	0.444			0.418	0.393

SAF, surface area fraction.

The regression parameter estimates and the percentage variation in coil mass permeability accounted for by the models are summarized in [table 4](#). For each PLS-PCR regression (Models B–E), analysis of variance indicated that the slope, which represents the transformed predictor along the principal component, was statistically significant and different from zero ($p < 0.0001$). The negative parameter estimates for the packing measures (PD, SAF_D and SAF_N) indicate that denser packing of the aneurysm will reduce coil mass permeability, which is in agreement with current knowledge. The positive parameter estimates for lacunarity (Λ_N and Λ_D) indicates that reducing gaps or improving spatial uniformity will decrease coil mass permeability. The linear regression model with PD as predictor (Model A) accounted for only 45% of the variation ($r^2 = 0.449$) in the measured permeability. Addition of coil type (C_T) in Model B better accounted for the variation in coil mass permeability. The C_T parameter estimates are denoted with a ‘±’ symbol, indicating that the estimate is positive for the GDC coil and negative for the Target coil. This implies that, when the aneurysms are coiled to the same PD, Target coils reduce permeability compared with GDC coils.

Of the three statistical models that considered measurements from the dome and neck sections, Model D incorporating measurements from the dome section (SAF_D and Λ_D) along with PD and C_T best explained the variation in permeability ($r^2 = 0.735$). This implies that the dome of the aneurysm needs to be densely and uniformly packed for reducing permeability. Interestingly, the percentage variation in permeability accounted for by Model E, which considered measurements from the neck section (SAF_N and Λ_N) along with PD and C_T , was not better than Model B which included PD and C_T . Plots of the measured versus the predicted permeability for four of the models (B–E) are shown in [figure 3](#) to aid in visual comparison. Model A was not included in the plot for clarity and because the coil type used for embolization is always known. The solid lines represent the linear fits to the data and are color coded to match the

symbols. Among the four models, the slope of Model D is closest to the identity line (dashes). [Table 5](#) summarizes the slope and intercept determined from the five models. Only Model D exhibited a slope with 95% CI spanning one and an intercept with 95% CI spanning zero, indicating good agreement between measured and predicted permeability.

DISCUSSION

The goal of endovascular embolization is to prevent aneurysm growth and rupture by decreasing the intra-aneurysmal blood flow.^{5–11} Using a patient-specific aneurysm model, the effect of PD, coil uniformity, and neck coverage on aneurysm permeability was observed. Although previous literature indicated that higher PDs result in lower intra-aneurysmal velocities and higher occurrences of stagnation,⁷ the study demonstrated that the permeability was influenced equally by uniformity and PD, as observed by similar loading factors or weights in [table 3](#). This implies that PD alone may not necessarily be an accurate predictor of aneurysm exclusion from coiling.

In previous studies, fluorescent microspheres were used to estimate regional perfusion and blood flow.^{21–22} The concept was extended to the investigation of aneurysm permeability in this study. The size of the microspheres (1 μm) used in this study was smaller than the pore size of the coiled aneurysm²³ and of similar size to platelets (3 μm). Therefore, the microspheres did not disrupt the coil mass or block the flow through the aneurysm. Although the described technique was useful for studying coil mass permeability, intra-aneurysmal thrombosis in response to the coil mass was not modeled in this in vitro experiment. Importantly, our study to measure coil mass permeability required an outlet, and therefore a pressure drop, across the aneurysm dome. The permeability measurements were therefore obtained in an idealized worst-case model system that is not representative of true intra-aneurysmal flow conditions. However, this design allows us to reproducibly study coil mass resistance to flow and to study the coil mass characteristics that are involved in reducing permeability.

The coil mass within the aneurysm acts as a porous medium, with voids in the coil mass creating channels for fluid travel. With a higher coil PD, the tortuosity—or weighted ratio of the length a fluid must travel to move across the media to the length of the media—increases. Because the porosity of a coiled aneurysm is imperfect, tortuosity cannot be assumed to be consistent throughout the aneurysm, supporting the idea that PD alone is not adequate for predicting treatment efficiency. When local porosity and tortuosity within an aneurysm (ie, of individual volumetric nodes) are considered, it becomes clear that the uniformity of the coil mass distribution will have an effect on the range of local porosities observed. Thus, in order to maximize

Table 4 Parameter estimates for the five statistical models

Statistical models	Intercept y_0	Coil type* C_T	Packing density PD	Dome section		Neck section		Variation in permeability accounted for (%)
				SAF _D	Λ_D	SAF _N	Λ_N	
Model A	1.292		−1.072					44.9
Model B	1.054	±0.042	−0.377					58.4
Model C	1.031	±0.021	−0.229	−0.115	0.525	−0.089	0.212	65.0
Model D	1.064	±0.029	−0.309	−0.153	0.615			73.5
Model E	1.039	±0.030	−0.296			−0.120	0.270	54.4

*± for coil type (C_T) parameter estimate denotes that C_T is positive for GDC coil and is negative for the Target coil.
SAF, surface area fraction.




Figure 3 Measured permeability plotted against the permeability predicted by the models. The solid lines are linear fit to the data and are color coded to match the symbols.

tortuosity—and maximize residence times as a result—the porosity range should be minimized and PD maximized. The range of local porosities can be minimized by taking care to ensure uniform distribution of the coil within the aneurysm is achieved.

Clinical implications of this study may include achieving a higher PD, and coil mass uniformity may be inversely related to the rate of aneurysm recurrence or coil compaction. A higher PD and uniform coil construct reduce aneurysm permeability, resulting in rapid exclusion of the aneurysm from the circulation. In this study a single operator coiled all the aneurysms, and thus dependence of homogeneity of the coil mass on the operator or technique cannot be assessed. However, two different coil types were used and we found that, compared with GDC coils, Target coils reduced coil mass permeability, primarily by improving packing measures (PD and SAF_D). The improved packing measures observed in the Target group may be expected because of the use of the longer framing coil, which is a limitation of this study. Since there is an upper bound to PD before coil herniation into the parent artery or catheter dislodgement occurs, our study indicates that designing coils to improve spatial distribution uniformity is a potential avenue for reducing aneurysm permeability.

Table 5 Slope and intercept values

		Model A	Model B	Model C	Model D	Model E
Intercept	Value	0.509	0.385	0.313	0.232	0.407
	SE	0.123	0.122	0.117	0.114	0.126
	95% LCL	0.245	0.123	0.061	-0.012	0.138
	95% UCL	0.773	0.647	0.565	0.476	0.676
Slope	Value	0.449	0.584	0.664	0.751	0.564
	SE	0.134	0.132	0.127	0.123	0.136
	95% LCL	0.164	0.301	0.392	0.488	0.273
	95% UCL	0.735	0.866	0.936	1.014	0.854

LCL, lower confidence limit; UCL, upper confidence limit.

This study presents a bench-top method to rigorously study coil uniformity and its impact on permeability. The selected aneurysm dimensions were based on our clinical database, which shows a mean diameter for coiled aneurysms of approximately 5 mm. This relatively small diameter is probably the result of numerous small ruptured aneurysms that are coiled as well as a shift for larger aneurysms to be treated with flow diverters. Future work should consider larger aneurysm models as well as other commercially available coils being specifically designed for large aneurysms.²⁴ Also under development are methods to reduce coil mass artifacts on 3D imaging.^{25 26} These techniques may ultimately lead to assessment of coil mass distribution using minimally invasive or non-invasive imaging modalities. However, until these developments are realized, bench-top experimentation can serve to provide feedback into the coil development process to produce coil technology that considers spatial distribution as a design input.

CONCLUSIONS

The results of this study reinforce the prior knowledge that PD influences coil mass permeability, with higher PD resulting in lower permeability. Importantly, the study showed that achieving spatially uniform distribution of coil mass with minimal gaps in the aneurysm dome is equally important in reducing coil mass permeability. A model incorporating PD, SAF of the dome section, and a spatial heterogeneity measure (lacunarity) of the dome section predicted coil mass permeability better than a model with PD alone. This study shows that uniformity and packing measures collectively and equally influence permeability.

Contributors J-YC, SV, AKW, SLC and MJG: conception and design. CB and SC: literature research. J-YC, SV, CB and MJG: statistical analysis and drafting the manuscript. All authors were involved in acquisition of data, analysis and interpretation of data, critical revision of the manuscript and approval of the final manuscript.

Funding This work was supported by Stryker Neurovascular. The content is solely the responsibility of the authors and does not represent the official views of Stryker. Abridged contents of this work have been accepted as a conference abstract at the 7th World Congress of Biomechanics.

Competing interests None.

Provenance and peer review Not commissioned; externally peer reviewed.

Open Access This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>

REFERENCES

- 1 Molyneux A, Kerr R, Stratton I, *et al.* International Subarachnoid Aneurysm Trial (ISAT) of neurosurgical clipping versus endovascular coiling in 2143 patients with ruptured intracranial aneurysms: a randomised trial. *Lancet* 2002;360:1267–74.
- 2 Molyneux AJ, Kerr RS, Yu LM, *et al.* International subarachnoid aneurysm trial (ISAT) of neurosurgical clipping versus endovascular coiling in 2143 patients with ruptured intracranial aneurysms: a randomised comparison of effects on survival, dependency, seizures, rebleeding, subgroups, and aneurysm occlusion. *Lancet* 2005;366:809–17.
- 3 Kole MK, Pelz DM, Kalapos P, *et al.* Endovascular coil embolization of intracranial aneurysms: important factors related to rates and outcomes of incomplete occlusion. *J Neurosurg* 2005;102:607–15.
- 4 Sluzewski M, van Rooij WJ, Slob MJ, *et al.* Relation between aneurysm volume, packing, and compaction in 145 cerebral aneurysms treated with coils. *Radiology* 2004;231:653–8.
- 5 Cha KS, Balaras E, Lieber BB, *et al.* Modeling the interaction of coils with the local blood flow after coil embolization of intracranial aneurysms. *J Biomech Eng* 2007;129:873–79.
- 6 Babiker MH, Gonzalez LF, Albuquerque F, *et al.* Quantitative effects of coil packing density on cerebral aneurysm fluid dynamics: an in vitro steady flow study. *Ann Biomed Eng* 2010;38:2293–301.

- 7 Otani T, Nakamura M, Fujinaka T, *et al.* Computational fluid dynamics of blood flow in coil-embolized aneurysms: effect of packing density on flow stagnation in an idealized geometry. *Med Biol Eng Comput* 2013;51:901–10.
- 8 Morales HG, Kim M, Vivas EE, *et al.* How do coil configuration and packing density influence intra-aneurysmal hemodynamics? *AJNR Am J Neuroradiol* 2011;32:1935–41.
- 9 Mehra M, Hurley MC, Gounis MJ, *et al.* The impact of coil shape design on angiographic occlusion, packing density and coil mass uniformity in aneurysm embolization: an in vitro study. *J Neurointerv Surg* 2011;3:131–6.
- 10 Chueh JY, Wakhloo AK, Gounis MJ. Neurovascular modeling: small-batch manufacturing of silicone vascular replicas. *AJNR Am J Neuroradiol* 2009;30:1159–64.
- 11 Plotin M, Iijima A, Wada H, *et al.* Increasing the packing of small aneurysms with complex-shaped coils: an in vitro study. *AJNR Am J Neuroradiol* 2003;24:1446–8.
- 12 Gunning PS, Saikrishnan N, McNamara LM, *et al.* An in vitro evaluation of the impact of eccentric deployment on transcatheter aortic valve hemodynamics. *Ann Biomed Eng* 2014;42:1195–206.
- 13 Lieber BB, Livescu V, Hopkins LN, *et al.* Particle image velocimetry assessment of stent design influence on intra-aneurysmal flow. *Ann Biomed Eng* 2002;30:768–77.
- 14 Serrat MA. Measuring bone blood supply in mice using fluorescent microspheres. *Nat Protoc* 2009;4:1779–58.
- 15 Mandelbrot BB. *The fractal geometry of nature*. New York: W H Freeman and Company, 1982.
- 16 N'Diaye M, Degeratu C, Bouler JM, *et al.* Biomaterial porosity determined by fractal dimensions, succolarity and lacunarity on microcomputed tomographic images. *Mater Sci Eng C Mater Biol Appl* 2013;33:2025–30.
- 17 Dominiotto M, Lehmann S, Keist R, *et al.* Pattern analysis accounts for heterogeneity observed in MRI studies of tumor angiogenesis. *Magn Reson Med* 2013;70:1481–90.
- 18 Kinney JP, Spacek J, Bartol TM, *et al.* Extracellular sheets and tunnels modulate glutamate diffusion in hippocampal neuropil. *J Comp Neurol* 2013;521:448–64.
- 19 Guidolin D, Zunarelli E, Genedani S, *et al.* Opposite patterns of age-associated changes in neurons and glial cells of the thalamus of human brain. *Neurobiol Aging* 2008;29:926–36.
- 20 Jayasuriya SA, Liew AW, Law NF. Brain symmetry plane detection based on fractal analysis. *Comput Med Imaging Graph* 2013;37:568–80.
- 21 Bartoli CR, Godleski JJ. Blood flow in the foreign-body capsules surrounding surgically implanted subcutaneous devices. *J Surg Res* 2010;158:147–54.
- 22 Glenny RW, Bernard S, Brinkley M. Validation of fluorescent-labeled microspheres for measurement of regional organ perfusion. *J Appl Physiol* 1993;74:2585–97.
- 23 Sadasivan C, Brownstein J, Patel B, *et al.* In vitro quantification of the size distribution of intrasaccular voids left after endovascular coiling of cerebral aneurysms. *Cardiovasc Eng Technol* 2013;4:63–74.
- 24 Mascitelli JR, Polykarpou MF, Patel AA, *et al.* Initial experience with Penumbra coil 400 versus standard coils in embolization of cerebral aneurysms: a retrospective review. *J Neurointerv Surg* 2013;5:573–6.
- 25 Prell D, Kyriakou Y, Struffert T, *et al.* Metal artifact reduction for clipping and coiling in interventional C-arm CT. *AJNR Am J Neuroradiol* 2010;31:634–9.
- 26 van der Bom IM, Hou SY, Puri AS, *et al.* Reduction of coil mass artifacts in high-resolution flat detector conebeam CT of cerebral stent-assisted coiling. *AJNR Am J Neuroradiol* 2013;34:2163–70.





Aneurysm permeability following coil embolization: packing density and coil distribution

Ju-Yu Chueh, Srinivasan Vedantham, Ajay K Wakhloo, Sarena L Carniato, Ajit S Puri, Conrad Bzura, Spencer Coffin, Alexei A Bogdanov, Jr and Matthew J Gounis

J NeuroIntervent Surg 2015 7: 676-681 originally published online July 16, 2014

doi: 10.1136/neurintsurg-2014-011289

Updated information and services can be found at:
<http://jniss.bmj.com/content/7/9/676>

These include:

References

This article cites 25 articles, 8 of which you can access for free at:
<http://jniss.bmj.com/content/7/9/676#BIBL>

Open Access

This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>

Email alerting service

Receive free email alerts when new articles cite this article. Sign up in the box at the top right corner of the online article.

Topic Collections

Articles on similar topics can be found in the following collections

[Basic science](#) (66)
[Editor's choice](#) (52)
[Open access](#) (44)

Notes

To request permissions go to:
<http://group.bmj.com/group/rights-licensing/permissions>

To order reprints go to:
<http://journals.bmj.com/cgi/reprintform>

To subscribe to BMJ go to:
<http://group.bmj.com/subscribe/>