

May 16th, 1:45 PM

Affordable Pressure Ulcer Prevention: Device, System and Algorithm

John McNeill

Worcester Polytechnic Institute

Yitzhak Mendelson

Worcester Polytechnic Institute

Raymond Dunn

University of Massachusetts Medical School

See next page for additional authors

Follow this and additional works at: http://escholarship.umassmed.edu/cts_retreat



Part of the [Biomedical Devices and Instrumentation Commons](#), and the [Translational Medical Research Commons](#)

McNeill, John; Mendelson, Yitzhak; Dunn, Raymond; Huang, Xinming; Sen, Devdip; and Hickle, Kelli, "Affordable Pressure Ulcer Prevention: Device, System and Algorithm" (2017). *UMass Center for Clinical and Translational Science Research Retreat*. 52.
http://escholarship.umassmed.edu/cts_retreat/2017/posters/52

This material is brought to you by eScholarship@UMMS. It has been accepted for inclusion in UMass Center for Clinical and Translational Science Research Retreat by an authorized administrator of eScholarship@UMMS. For more information, please contact Lisa.Palmer@umassmed.edu.

Presenter Information

John McNeill, Yitzhak Mendelson, Raymond Dunn, Xinming Huang, Devdip Sen, and Kelli Hickle

Keywords

pressure ulcers, prevention, sensors

Creative Commons License

This work is licensed under a [Creative Commons Attribution-Noncommercial-Share Alike 3.0 License](https://creativecommons.org/licenses/by-nc-sa/3.0/).

AFFORDABLE PRESSURE ULCER PREVENTION: DEVICE, SYSTEM AND ALGORITHM

John McNeill, PhD¹, Yitzhak Mendelson, PhD¹, Raymond Dunn, MD², Xinming Huang, PhD¹, Devdip Sen¹ Kelli Hickie MD²

¹Worcester Polytechnic Institute, Worcester, MA; ²University of Massachusetts Medical School

The incidence and prevalence of wound problems in an aging and longer-lived population continue to increase substantially. Pressure ulcers (bedsores) are painful, increase risk for secondary infection, and add \$ 11 billion annually to health care costs in the US. When a sufficient fraction of the patient's weight is supported in a region with a bony prominence, the resulting localized concentration of external pressure reduces the cross-sectional area of blood vessels, restricting blood flow and limiting oxygen supply to the at-risk tissue. If this external contact pressure is maintained for sufficient time, the lack of oxygen leads to tissue death and formation of a pressure ulcer. Very few effective ulcer prevention devices exist and those available have focused almost exclusively on some form of off-loading (specialized and costly beds and wheelchair cushions) designed to equalize pressure distribution on at-risk tissue areas. Furthermore, wheelchair cushions have no applicability to other body areas (legs, heels, etc.) at risk for pressure injury. Despite extensive work in this area of pressure ulcer prevention, a survey of available options shows that there is no compact, low-cost solution that is workable in a hospital setting, for patients confined to bed at home, or for those with limited mobility in wheelchairs.

In this work we present a wireless sensor patch to be placed on known at-risk sites on the patient's skin, directly measuring local contact pressure and temperature, and communicating with a base station (in a hospital setting) or smartphone (for home care). Alerting a patient or caregiver to a potentially harmful level of pressure allows early intervention (only when necessary) to prevent pressure ulcer formation, easing workload on caregivers and enabling more independence for mobile at-risk patients.

Contact:

Devdip Sen

Worcester Polytechnic Institute

dsen@wpi.edu